

# Pauli Weighing Systems

Specialist in Weighing Solutions



Company:

CERT No:

Site:

ID No:

<b>Equipment Details</b>	<b>Job No: AF290413-1</b>
Description: Type/Model: Serial No: Capacity: Increment: Location Building:	Comments: Recalibrated.

<b>Calibration Details</b>		<b>Error Allowance</b>	<b>± 0.0015g</b>
<b>Calibration Equipment:</b>	<b>Certificate Number:</b>	<b>Exp Date:</b>	<b>Uncertainty</b> ± 0.0001%

Cumulative Uncertainty: ± 0.0001%

Next Calibration Due:

**Equipment Display Unit 'g'**

As Found				Definitive			
Weight	Reading	Error	Pass / Fail	Weight	Reading	Error	Pass / Fail
0	0.0000	0	Pass	0	0.0000	0	Pass
20	20.0003	0.0003	Pass	20	20.0003	0.0003	Pass
50	49.9997	-0.0003	Pass	50	50.0003	0.0003	Pass
100	99.9986	-0.0014	Pass	100	100.0004	0.0004	Pass
150	149.9975	-0.0025	Fail	150	150.0005	0.0005	Pass
200	199.9960	-0.0040	Fail	200	200.0005	0.0005	Pass
150	149.9975	-0.0025	Fail	150	150.0010	0.0010	Pass
100	99.9986	-0.0014	Pass	100	100.0006	0.0006	Pass
50	49.9991	-0.0009	Pass	50	50.0003	0.0003	Pass
20	20.0006	0.0006	Pass	20	20.0005	0.0005	Pass
0	0.0000	0	Pass	0	0.0000	0	Pass

**Eccentric Load Test Error Allowance 'e' 50g**

As Found		Definitive	
Reading: 49.9996	Reading: 49.9997	Reading: 50.0003	Reading: 49.9998
Error: -0.0004	Error: -0.0003	Error: 0.0003	Error: -0.0002
Reading: 50.0001	Reading: 49.9997	Reading: 49.9999	Reading: 50.0004
Error: 0.0001	Error: -0.0003	Error: -0.0001	Error: 0.0004

Pauli Weighing Systems Ltd Registered in England – Registration Number: 8545629

Repeatability	Applied	Zero	Load	Zero	Load	Zero	Load
	50g	0.0000	50.0003	0.0000	50.0003	0.0000	49.9999

Date:

Calibrated by: